

# Loose screw?

These are the suitable tools!



Instruments remain  
in the tray  
during cleaning



## A good range\*

The basic kit offers you a selection of the most common tools for daily use.

Allen (inch)	0.03	0.05		
Allen (mm)	0.9	1.0	1.2	1.8
Torx (mm)	SKY	5.5	6.0	
Slotted (mm)	1.6	2.0		

\*contains contra-angle instruments in the long version

## Tighten with the Full Range Driver Kit now!

**Universal** | can be used for prosthetic screws of the major manufacturers on the market

**Hygienic** | validated sterilization and reprocessing process

**Convenient** | during cleaning the instruments remain in the tray in the thermal disinfecter




Full Range Driver Kit  
with contra-angle instruments  
**REF 580DRIVE**




Full Range Driver Kit  
Empty storage box without instruments  
**REF 580DRIVO**




SKY  
Torque Wrench Pro Set  
(ratchet +adapter)  
**REF SKYTWSET**




CPS Cordless  
Prosthodontic  
Screwdriver  
**REF 580CPS39**




SKY  
laboratory handle  
**REF SKY-SD80**

**In addition to the standard equipment of the Full Range Driver Kit, the following products can be ordered optionally:**

Qty	REF	Designation
	SKYTWPRO	SKY Torque Wrench Pro
	SKYTWCON	SKY Connector Pro
	SKY-DV12	SKY drill extension
	SKY-SD22	SKY prosthetic key for contra-angle, short
	310W0K01	Screwdriver 1 short, Torx 6, contra-angle
	310W0K02	Screwdriver 2 short, slotted 1.6, contra-angle
	310W0K03	Screwdriver 3 short, slotted 2, contra-angle
	310W0K05	Screwdriver 5 short, Allen 0.05", contra-angle
	310W0K06	Screwdriver 6 short, Allen 0.9, contra-angle
	310W0K07	Screwdriver 7 short, Allen 1.0, contra-angle
	310W0K08	Screwdriver 8 short, Allen 1.2, contra-angle
	310W0K09	Screwdriver 9 short, Allen 1.8, contra-angle
	310W0110	Screwdriver 10 long, Allen 2.5, contra-angle
	310W1011	Screwdriver 11 long, square 1.3, contra-angle
	310W0K11	Screwdriver 11 short, square 1.3, contra-angle
	310W0K12	Screwdriver 12 short, Torx 5.5, contra-angle
	33001155	Torque ratchet 10-40 Ncm
	58001168	Ratchet adapter

Stamp

Date ..... Signature ..... Order

Date ..... Signature .....  
Declaration of consent in accordance with our data privacy statement No X.



009924GB-20181030 Mistake and subject to change reserved